



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Test | Target/UOM | Range | Result |
|--|---|-----------------|------------|
| Packaged Product: EQS050F | | | |
| | HYDROCHLORIC ACID 0.02 NORMAL (SPECIAL) | | |
| PREPARED TO FORMULATION ON FILE | YES | | YES |
| Sample ID | | | |
| Tested 05/23/24 | | | |
| Tester MMAT | | | |
| Recorded 05/23/24 | | | |
| SAMPLE COLLECTION & HANDLING PER LAB040 | YES | | YES |
| Sample ID | | | |
| Tested 05/23/24 | | | |
| Tester MMAT | | | |
| Recorded 05/23/24 | | | |
| DATE OF SAMPLING FOR TEST PURPOSES | MM/DD/YYYY | | 05/23/2024 |
| Sample ID | | | |
| Tested 05/23/24 | | | |
| Tester MMAT | | | |
| Recorded 05/23/24 | | | |
| AMBIENT TEMPERATURE DURING SAMPLING DEGC | 25.0 DEG C | 10.0 - 35.0 | 21.0 |
| Sample ID | | | |
| Tested 05/23/24 | | | |
| Tester MMAT | | | |
| Recorded 05/23/24 | | | |
| DATE OF TESTING OF SAMPLE | MM/DD/YYYY MM/DD/YYYY | | 05/23/2024 |
| Sample ID | | | |
| Tested 05/23/24 | | | |
| Tester MMAT | | | |
| Recorded 05/23/24 | | | |
| AMBIENT TEST TEMPERATURE DEG C | 20.0 DEG C | 10.0 - 30.0 | 21.0 |
| Sample ID | | | |
| Tested 05/23/24 | | | |
| Tester MMAT | | | |
| Recorded 05/23/24 | | | |
| Actual Normality | 0.0200 EQ/L | 0.0195 - 0.0205 | 0.0198 |
| Sample ID | | | |
| Tested 05/23/24 | | | |
| Tester MMAT | | | |
| Recorded 05/23/24 | | | |
| APPEARANCE AND COLOR | CLEAR LIQUID APPEARANCE | | PASS |
| Sample ID | | | |
| Tested 05/23/24 | | | |
| Tester MMAT | | | |
| Recorded 05/23/24 | | | |



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| INSTRUMENTS USED DURING PREPARATION | INSTRUMENT | P-58,VF2-A6,M55,BID02699,PI DP123 |
|--|------------|--------------------------------------|
| Sample ID Tested 05/23/24 Tester MMAT Recorded 05/23/24 | | |
| Lot # 4052158 Made 05/23/24 | | |

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.


Quality Manager