

## Aqua Solutions, Inc.



P.O. Box 70 Deer Park, TX 77536 Ph (281-479-2569)

SPECIFICATION SHEET	
NITRIC ACID 0.01 NORMAL	

Product Code	6459
Product Name	Nitric Acid 0.01 Normal
N.I.S.T. Traceable	Yes
Actual Normality	0.0100 +/- 0.0005 Normal @ 20° C
Prepared to Formulation on File	Yes
Appearance and Color	Clear water white liquid
Country of Origin	USA
Harmonization Code	2808.00.0000

The weights and /or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All Volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

PREPARATION DATE (MM/DD/YY)		
Original 0.0	04/30/2018: STN	
Revision		

PLEASE REFER TO THE MATERIAL SAFETY DATA SHEET FOR HEALTH, SAFETY, AND REGULATORY INFORMATION ON THIS PRODUCT.